



### **CLEANING OF NEBULIZER:**

After each use, unscrew nebulizer into parts, wash in warm soapy water, and rinse well. Leave in a clean area to dry.

### **NOTE:**

Do not towel dry nebulizer parts; this could cause contamination.

### **CLEANING ACCESSORIES:**

After each use rinse and dry mouthpiece. Never Immerse tubing in water. Single patient use for one month.